



This report contains two sections: Chain of Custody and Report.

| Customer ID | Location | Volume (L) | Fibers | Fields | F/mm ² | F/cc | Comment |
|-------------|-------------------------|------------|--------|--------|-------------------|---------|--|
| 1 | ICA/ North | 1020 | 5.0 | 100 | < 7.0 | < 0.010 | |
| 2 | ICA/ West | 1050 | 5.5 | 100 | 7.0 | < 0.010 | |
| 3 | OCA/Decon | 1000 | 5.0 | 100 | < 7.0 | < 0.010 | |
| 4 | OCA | 1020 | 47.0 | 100 | 59.9 | 0.023 | |
| 5 | OCA | 1020 | 5.0 | 100 | < 7.0 | < 0.010 | |
| 6 | ICA/Area North | 900 | 23.0 | 100 | 29.3 | 0.013 | |
| 7 | ICA/Area East | 1200 | 0.0 | 100 | < 7.0 | < 0.010 | |
| 8 | ICA | 1200 | 12.0 | 100 | 15.3 | < 0.010 | |
| 9 | West | 480 | 5.5 | 100 | 7.0 | < 0.010 | |
| 10 | OCA Entrance | 450 | 100.0 | 35 | 364.0 | 0.311 | Sample result may be biased due to dust loading. |
| 11 | OCA/East | 1200 | 0.0 | 100 | < 7.0 | < 0.010 | |
| 12 | BFB Open Blank | 0 | 0.0 | 100 | < 7.0 | N/A | |
| 14 | OCA/AFD Northeast | 1050 | 100.0 | 20 | 636.9 | 0.234 | Sample result may be biased due to dust loading. |
| 15 | Wall | 1200 | N/A | N/A | N/A | N/A | Void: Overloaded with particulate. |
| 16 | ICA/Area North | 1700 | N/A | N/A | N/A | N/A | Void: Overloaded with particulate. |
| 17 | ICA/Area West | 700 | 1.5 | 100 | < 7.0 | < 0.010 | |
| 18 | Environmental West Side | 1200 | N/A | N/A | N/A | N/A | Void: Filter wet. |
| 19 | West Side | 1200 | N/A | N/A | N/A | N/A | Void: Overloaded with fibers. |
| 20 | Exhaust | 1200 | N/A | N/A | N/A | N/A | Void: Cassette manufactured incorrectly. |

Tech. Notes:

Submitted By: John Doe | via: U.S. Mail | Submittal Date: 5/1/2017 1:47:53 PM | Sample Date: 5/16/2017 | Report Date: 9/21/2017 | Lab Job No.: 17-1267 | Analysis Date: 9/21/2017 | Technician: Sally Scientist

Results apply only to samples tested. Results may not be reproduced or reported except in full without prior written approval of Moldlab. All samples were received in acceptable condition unless noted in the Tech Notes section. All calculations are corrected for blanks (if provided). Microscope Field Area: 0.00785 mm².

TDSHS License No. 300431



2501 Mayes Rd #110
Carrollton, Texas 75006
P - (972) 820-9373
Toll Free (866) 416-6653
Website - www.moldlab.com